

Manual Supplement

Manual Title: Multipoint Power Recorder Installation Guide
Part Number: 2010675 Supplement Issue: 1
Print Date: December 2003 Issue Date: 11/05
Revision/Date: 3, 1/04 Page Count: 1

This supplement contains information necessary to ensure the accuracy of the above manual.



Change #1

On page 39, under **Voltage**, replace both Voltage measurement categories with the following:

Measurement category: 600 V CAT III Pollution Degree 2

Under **Current**, add the following:

Current Range (with external Current Transformer):
10 mA to 5000 A.

On page 40, under **Environmental Requirements**, change the following:

From: For use indoors or in suitable, protective enclosure outdoors.

To: For indoor use only.

From: Operating: 0 °C to 50 °C

To: Operating: 0 °C to 40 °C

From: Humidity: less than 90 % non-condensing

To: Humidity: Maximum relative humidity 80 % up to 31 °C, decreasing linearly to 50 % relative humidity at 40 °C.